

MADE IN
EUROPE



This product has been developed and manufactured using both raw materials and with production in Europe.

Choose quality, save money, and protect the environment.


(UK) User Manual

01_20260414_UK

Master Glide Transfer



MC 004-1012



Regulation MDR 2017/745 for medical devices class 1



ISO 12 31 03 18

The Original Nordic
Human Handling Concept



Scan Me

Developed &
Manufactured by:



REF (Article no.)	HMI no.	Name - Size	UDI-DI/GTIN
MC 004-1012	129464	Master Glide Transfer - L	05744001972649

Thank you for choosing a product from master care A/S.

Important information before using the product(s):

We recommend reading the user manual thoroughly before using the product (-s).

The information in this user manual applies to the shown products provided that the products have not been modified or customized and that the products have been used as recommended in this user manual.

If the customer changes, repairs or uses the product in other combinations than recommended by master care A/S the CE-approval no longer applies and the guarantee becomes void.

In case of doubt, contact master care A/S.

MDR (EU) 2017/745:

Products in this user manual is CE-approved and follows the harmonized standards. This means that the products are in compliance with the essential health and safety directives. The product(-s) comply with the standards in the Medical Device Regulation MDR (EU) 2017/745 – Class 1.

The intended use of the product/-s:

The product can be used for lying transfers of users between two surfaces. It is recommended to use sliding material and/or a draw sheet.

Proper use of the product:

Place the board halfway under the user after which the user slides over to the receiving bed. Nylon should be used if the transfer is considered difficult/hard or for users who are at risk of sustaining shear injuries.

Indication:

The board can be used in connection with horizontal transfers, for example between a stretcher/operating table and a bed.

Contraindication:

Due to the board's narrow width, it should not be used for bariatric patients.

No other contraindications for the use of the product are known.

Warning:

The board should not be used over longer unsupported distances than that indicated on the board.

Risk assessment:

Always make a risk assessment and a functional assessment of the user prior to using the product. This way using the product for transferring or moving a user, is ergonomic and legally sound.

Knowledge of the product/transferring:

It is important that the user of the product has been properly instructed in the use of the product and has read the user manual thoroughly prior to using the product.

master care A/S provide education in the use of products and on our website all user manuals can be found.

Date of manufacture and Life span:

The product's production date can be found on the product engraving under 'LOT' and is displayed as YYMMDD - e.g. 251020, which refers to the product being produced on the 20th of September 2025.

The expected life span of the products with daily wear and tear is averagely 2-5 years. The products life span depends on the use and user's load.

Inspection:

Make visual inspections frequently and preferably before every use. Products with insufficient friction reduction, broken material or other deficiencies must be discarded. This way you will always keep a sound and healthy work environment for both helper and user.

Guarantee:

master care A/S provide a 2-year guarantee covering material defects and manufacturing defects if the product has been used correctly.

Unintended events and injuries when misusing the products:

Unintended events must always be reported to the correct national authority and to master care A/S or the distributor.

master care A/S takes no responsibility for personal injuries or product damages if the product(-s) have been used in contravention of the by the manufacturer recommended intended purpose or proper use of the product.



User Manual - Master Glide Transfer

Mount and use of Master Glide Transfer Board:



1. Tip the user slightly towards side lying position. Use a drawsheet if possible.



2. Place Master Glide Transfer Board under the head and pelvis. Make sure that the head is fully supported by the board.



3. Use sliding material between the sheet and the board if the user is heavy, or the transfer is otherwise hard.



4. Adjust the receiving bed 2-5 cm lower than the primary bed. Remember to lock all relevant breaks.



5. The user bends the legs. The helpers respectfully pull and pushes the use at a calm pace towards the receiving bed.



6. Remove the board and pull out the sliding material. Straighten out the drawsheet or remove it.

User Manual - Master Glide Transfer

Technical Specifications

REF - Product Name:	Size:	Handles:	Max. user weight:	Product weight:
MC 004-1012 Master Glide Transfer - L	L 120 x B 60 x H 0,5 cm	10	160 kg	3,32 kg

Materials:
PE

The products meet the criteria in the following standards:

REACH

RoSH

DODD Frank

WARNINGS



Never leave the user on the product when lying on the edge of the bed.



Never leave the product on the floor.



Max 20 cm between the support surfaces during transfer.



Make sure that all relevant brakes are activated before transferring the user.



CE-certification:

This product is CE-approved and follows the harmonized standards. This means that the products are in compliance with the essential health and safety directives. The product(-s) comply with the standards in the Medical Device Regulation (EU) 2017/745 – class 1.



User weight (SWL - safe working load):

Recommended max. user weight: 160 kg



Surface cleaning:

The surface of this product can be cleaned with a wet cloth and soap.



Detergents pH level:

This product tolerates detergents / disinfectants with a pH level between 5-9.



Disinfectants:

This product tolerates alcoholic disinfectants with a pH level between 5-9.



Maintenance:

Make visual inspections frequently and preferably before every use. Pay attention to scratches and tears to the surface. Products with high friction or other deficiencies must be discarded.



User Manual:

Always read the User Manual thoroughly before using the product.



Disposal:

Always make sure to follow local guidelines when disposing or discarding the product.

For more information, videos
Quick Guide etc., visit
www.master-care.dk

