



Human Handling Concept



EU DECLARATION OF CONFORMITY

Name and address of the firm: master care A/S
Sofienlystvej 3, 8340 Malling, Denmark

Single Registration Number (SRN): DK-MF-000011791

We, as the manufacturer, declare that we have issued this EU declaration of conformity under our sole responsibility and that the products listed below, all class 1 according to annex VIII of MDR 2017/745 – rule 1, meets all the provisions of the MDR 2017/745 which apply to conformity assessment procedure MDR Art. 52 & Annex II, III (Class I).

Classification: Medical Device, Classification 1.

Product name: Master Glide Sif

Basic-UDI-DI: 57440019763331GY

Article no.: MC 001-7060

Intended use:

1. The product is placed in pairs underneath the user's thighs to reduce friction when the user moves or is moved backwards on the seat.
2. Place the pair under the user's thighs. Slide the leg strap of the sling between the plates to minimize the risk of shear injuries.

Declaration's validity:

5 years after signing or changes to the technical documentation.

Date of Issue and place: Malling, 15.01.2025

Name and function:


mastercare
HUMAN HANDLING SYSTEM

master care A/S • Sofienlystvej 3 • DK-8340 Malling
Tlf. 80 83 83 85 • info@master-care.dk
www.mastercare.dk

Vidar Geitheim, CEO