



## Human Handling Concept



### EU DECLARATION OF CONFORMITY

Name and address of the firm: master care A/S  
Sofienlystvej 3, 8340 Malling, Denmark

Single Registration Number (SRN): DK-MF-000011791

We, as the manufacturer, declare that we have issued this EU declaration of conformity under our sole responsibility and that the products listed below, all class 1 according to annex VIII of MDR 2017/745 – rule 1, meets all the provisions of the MDR 2017/745 which apply to conformity assessment procedure MDR Art. 52 & Annex II, III (Class I).

Classification: Medical Device, Classification 1.

Product name: Master Glide Thor

Basic-UDI-DI: 57440019763336HA

Article no.: MC 001-7061

#### Intended use:

1. The intended use is to create a bridge between two surfaces during lying transfers of a user with decreased physical function. Can be used with nylon or/and with a Disposable Cover.
2. The product is used to make it easy and more comfortable to place and remove a sling. It reduces the risk of shear injuries and skin tearing on back.

#### Declaration's validity:

5 years after signing or changes to the technical documentation.

Date of Issue and place: Malling, 15.01.2025

Name and function:

 **master care**  
HUMAN HANDLING SYSTEM  
master care A/S • Sofienlystvej 3 • DK-8340 Malling  
Tlf. 86 98 8170 • info@master-care.dk  
www.master-care.dk

Vidar Geitheim, CEO