



## Human Handling Concept



### EU DECLARATION OF CONFORMITY

Name and address of the firm: master care A/S,  
Sofienlystvej 3, 8340 Malling, Denmark

Single Registration Number (SRN): DK-MF-000011791

We, as the manufacturer, declare that we have issued this EU declaration of conformity under our sole responsibility and that the products listed below, all class 1 according to annex VIII of MDR 2017/745 – rule 1, meets all the provisions of the MDR 2017/745 which apply to conformity assessment procedure MDR Art. 52 & Annex II, III (Class I).

Classification: Medical Device, Classification 1.

Product name: Master Glide Junior, Senior, Flexi, Viking, Multi,  
Bifrost, Amput.

Basic-UDI-DI: 57440019763252QL

Article no.:

MC 001-6000, MC 001-6010, MC 001-6020, MC 001-7000, MC 001-7010, MC 001-7020, MC 001-7026,  
MC 001-7027, MC 001-7030, MC 001-7035, MC 001-7036, MC 001-7037, MC 001-7040, MC 001-7055,  
MC 001-7056.

Intended use:

The intended use is to create a bridge between to surfaces during seated or semi-standing transfer between a chair and bed. For users who have difficulty letting go of the surface during transfer due to decreased physical or cognitive functions.

Double leg amputee users should be cautious not to tilt the board during transfer.

Declaration's validity:

5 years after signing or changes to the technical documentation.



master care A/S • Sofienlystvej 3 • DK-8340 Malling  
Tlf. 86 98 85 00 • info@master-care.dk  
www.master-care.dk

Date of Issue and place: Malling, 15.01.2025

Name and function:

Vidar Geitheim, CEO